IISER/PUR/PT/1/15

INDEX

- 1. Multi Source Thermal Evaporator
- Page No. 1
- 2. Liquid Helium Plant Page No. 4
- 3. Atomic Force Microscope/ Scanning Tunneling Microscope
 - Page No. 9

Sd/-



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH THIRUVANANTHAPURAM [IISERTVM]

PH.-0471 2597454, FAX: 0471-2597427 CET CAMPUS, ENGINEERING COLLEGE. P.O THIRUVANANTHAPURAM 695016, KERALA, INDIA

EMAIL: <u>purchasestores@iisertvm.ac.in</u>
IISER/PUR/PT/1/15

6th May 2015

INVITATION FOR BIDS

Sealed and superscribed **Two-Part** bids are invited for supply of the following equipments.

I. MULTI-SOURCE THERMAL EVAPORATOR - 1 NO

1. TENDER NO. : IISER/PUR/4173/14

2. Due Date : 2nd June 2015 [4 PM]

3. Date of Opening of : 3rd June 2015 [3 PM]

TECHNICAL BID

4. DATE OF OPENING OF PRICE: Will be notified later.

BID

5. Cost of Tender * : Rs. 500 + 4.04% VAT

6. EMD [BID SECURITY] ** : Rs. 1 Lakh

SPECIFICATIONS:

Application: Thermal evaporator for coating (i) Metals: Au, AI, Ag, Ca, Cr, Mg, Pd etc + alloys (ii) Inorganic compounds: MoO3, LiF, ZnO, TiO₂, etc. upto 300 nm and with an accuracy of 0.1 nm.

Detailed System Configuration:

1. Chamber Design

- a) Chamber to be manufactured of non-magnetic SS
- b) The chamber size should be not less than 14" wide x 14" deep x 24" height (internal dimensions)
- c) Provision for backfilling the chamber with air/ dry nitrogen should be provided.
- d) Viewport with
- e) Dismountable chamber wall liners for cleaning deposited materials.
- f) Chamber stand
- 2. Vacuum Requirements and Pumping System
- A. Chamber base pressure ~ 1e-7 mbar assisted by a liquid nitrogen trap. Since this system will be used for ohmic contacts, clean vacuum is absolutely crucial. The vacuum during evaporation should be 5×10^{-6} mbar or better. Please provide necessary technical details.
- B. Air-cooled turbo molecular pump with pumping speed ~600It/sec. The backing pump should be a dry scroll pump.
- C. Under clean conditions the pump down from atmosphere to base pressure should be possible within 1 hours.
- D. A metal mesh guard should be provided over the entry to the pumping port of the turbo to ensure that small flakes of metal/substrate/screws etc do not drop into the turbo blades.
- E. A gate/plate valve should be provided at the base of the chamber that allows full isolation of the chamber under vacuum, for clean storage over idle periods.
- F. We prefer electronic switching from roughing mode to backing mode. But manual switching -by the user (opening appropriate valves after certain

pressures are reached) after certain pressures are reached would also be acceptable, if this means substantial reduction in cost.

Quartz Crystal Oscillator + Monitor/Controller Set:

A 5/6 MHz quartz crystal sensor head and monitor for in-situ thickness monitoring. Controller should be capable of storing multiple values of density and acoustic impedance in the memory for repeated use and have a thickness resolution of 0.1 nm. It should include all accessories (e.g. water cooling tubes, Oscillator + Cables + Spare crystals set). The position of the quartz crystal inside the chamber should be typically at the same level as the sample holder, but for controlling thickness and uniform coating a shutter need to be placed between the quartz crystal and substrate.

Source Specifications:

Design of the electrodes should be such that four boats/spirals are arranged radially outward from a central post. The wheel (base plate) should be rotatable from outside (manually) so that the desired boat can be brought into contact with one fixed electrical post/spring loaded brush at the circumference of the wheel. Manually operated shutter for covering the sample from the sources. It should be designed such that it may be stationed in 3 positions (i) fully open (ii) covering the sample but not the quartz crystal head and (iii) covering both, the sample and the quartz crystal head. Possibility of co deposition

Substrate Holder Specifications:

- a) Sample holder size 150 mm x150 mm (maximum) and typical size 10 mm x 10 mm.
- b) Typical substrate to source distance ~ 20 cm
- c) Arrangement for rotating the sample holder at adjustable speeds upto about 10 RPM, with a motor fully inside the chamber.
- d) Tilt will be fixed at the time of loading the sample and need not be adjustable from outside.
- e) Substrate Heating: 800 deg C maximum
- f) Note: The substrate heating and rota-tilt facilities may be supplied in two separate user interchangeable versions. Either may be used on user requirement.

Spare parts and consumable availability:

50 dimpled boats, 50 spirals, 50 baskets + crucibles and 10 Viton gaskets of each type used on the various ports of the system. The size and design of these boats/spirals/baskets should be discussed with the customer prior to design and supply.

Operating voltage and emergency shutdown:

The full unit should operate from single-phase Indian AC mains: 240VAC, 50 Hz, single phase.

Water Chiller: A compact water chiller capable of supporting up to two evaporators of the type supplied should be included.

General Requirements:

- a) System should be Clean Room compatible class 1000.
- b) Safety Interlocks for water supply, high voltage, vacuum.
- c) Utility module for connection to utilities like water with water distribution manifold, compressed air and power etc.
- d) A guarded switch should be provided on the front panel for emergency shutdown.
- e) Chamber design to be supplied, discussed and with customer before fabrication.
- f) The quoted price should include installation at customers' laboratory.

- g) The vendor must demonstrate compliance with tendered specifications at their factory before acceptance of delivery to IISER. The same must be redemonstrated upon installation at IISER.
- h) Extended warranty for 3 years -beyond the mandatory Optional Items:
 - a. Automatic control of thin film deposition -based on a thickness controller integrated with source power/shutter.
 - b. Additional Power Sources for co-evaporation from 2 sources.
 - c. Annual Maintenance contract.

II. <u>Liquid Helium</u> <u>Plant</u> - 1 No <u>SCHEDULE</u>

1) TENDER NO. : IISER/PUR/5096/14

2) DUE DATE : 2nd June 2015 [4 PM]

3) DATE OF OPENING OF : 3rd June 2015 [3 PM]

TECHNICAL BID

4) DATE OF OPENING OF PRICE : Will be notified later.

Bid

5) Cost of Tender # : Rs. 500 + 4.04% VAT

6) EMD [BID SECURITY] ** : Rs. 9 Lakhs

SPECIFICATIONS:

A) Required Liquefying capacity 10 to 15 liters per hour, with 2% air impurity in feed helium gas without liquid nitrogen pre-cooling, as measured by the method of level rise in the Mother Dewar at nearly atmospheric pressure. The system should operate with electrical power supply of 415 VAC + or - 10%, 3-phase, 50 Hz.

B) Cold Box:

- ➡ Integrated automatic helium re-purifier capable of handling up to 15% air impurity in feed gas.
- Liquid nitrogen pre-cooling facility. Specify the liquefying capacity with the liquid Nitrogen pre-cooling facility.
- Fully pre-programmed PLC Control panel with suitable operator panel.
- → Full automatic control for cool-down, operation and shut-off with adequate safety devices for un-attended operation.

C) Compressor:

- Rotary screw type recycle compressor, (offer for both air cooled & water-cooled version) what are the other types of compressor?
- → Oil removal system and Gas management panel.
- → Line drier unit capable of handling moisture contents in the feed helium gas containing up to 20% air, assuming atmospheric conditions of our Institute location in India.

D) Liquid Helium Dewar (Mother Dewar):

- → Liquid helium Dewar of 1000 liters capacity.
- Liquid level sensors and local digital indicator directly calibrated in liquid liters.
- → Equipped with mountings viz. Liquid withdrawal port, electrical pressure building system, compound manometer, safety valves, vent valves, vibration damper, rupture disc etc.,

E) Essential Accessories:

- 1) Pure gas buffer gas-vessel suitable for the above liquefier when operating with pre-cooling.
- 2) Remote co-axial delivery tube compatible with the Mother Dewar.
- 3) Supply of two numbers of Recovery helium gas balloon (capacity around 10 cubic meters, roof hanging type) with suitable relief valves and supports but without level sensors.
- 4) External cryogenic charcoal adsorber with accessories like heater, hose pipes, interconnecting pipes etc.
- 5) High pressure Helium-gas recovery gas-compressor (~30 cubic meters per hour capacity, 200 bar operating pressure, air cooled, silent version) with built-in purification system, electronic control with intermediate pressure/temperature monitoring along with the complete set of consumable spares.
- 6) Instrument air compressor with suitable inbuilt refrigerated drier unit.
- 7) Remote monitoring and control system on PC for remote on/off: process visualization with dynamic display, trend recording, display of control loop status and with remote diagnostics facility including necessary desktop with preloaded software, 22 Inch LED monitor, software package/ license etc.
- 8) Complete piping kit and associated components the installation of the system.
- 9) Maintenance toolkit should be provided for routine maintenance of the system.
- 10) Technical assistance/Supervision of installation, commissioning of the plant: If the installation of other components, which includes plant room layout, piping layout, welding, installation of interconnecting pipes, leak testing, fixing/clamping at site done by a third party.
- 11) Operator training for the liquefier at our site.
- 12) Final performance & acceptance test of helium liquefier as mentioned in the above offer at our site.
- 13) Operation and maintenance manual (both soft & hard copy)
- **F) Warranty:** 24 months from the date of commissioning.

G) Optional Accessories:

- (1) Installation of the liquefier and other components, which includes plant room layout, piping layout, welding, installation of interconnecting pipes, leak testing, fixing/clamping at site.
- (2) Inspection, testing and training of one Institute engineer at the suppliers' works during tests, performance and shop acceptance of the whole liquefier. The offer is understood to include all the expenses for inspection and training including round trip economy fare, stay etc. of the engineer.
- (3) Flexible transfer Siphon for liquid withdrawal from the mother dewar to the transport dewar supplied with suitable godard coupling matching the above liquid helium transport dewars.
- (4) External chiller unit of suitable capacity for the feed compressor.
- (5) Major maintenance kit for the compressor and the cold box (consumable & nonconsumable).
- (6) AMC for five years from the date of completion of standard warranty period.

H) Essential requirements for helium liquefier

- (1) The offered model and type of the liquefier should be highly energy efficient, less in maintenance cost, rugged and highly reliable in operation and control.
- (2) The offer must confirm that the liquefier is capable of unattended, fully automatic, foolproof and fail-safe operation and is equipped with adequate number of Annunciators, Alarms and Safety devices. It must also contain the description and operation of such automation systems and devices. 3 4
- (3) The liquefier must be operable on the electrical supply parameters available at the Institute as: 415 V AC plus or minus 10%, 3-phase, 50 Hz. The offer must certify that all the machinery, equipment, instruments and controls will keep working and will remain safe at these parameters of electrical supply
- (4) The offer must certify that, in case the electrical supply parameters change beyond the range given above, the system may shut-down but all its components will remain safe from any damage.
- (5) The offer must confirm that the liquefier operating procedure will allow transfer of liquid helium from the Mother Dewar into child dewars without adversely affecting the liquefaction process in any mode of operation
- (6) The offer must give performance data on guaranteed liquefying rates in various operating modes viz. (a) With and without liquid nitrogen precooling (b) With various purities of feed helium gas starting from 99.995% to 80%.
- (7) The offer must certify that the liquefier and its components will safely withstand the site conditions of our institute.
- (8) The offer must contain full details of the electric power requirements of all components, air and water requirement, floor preparation if needed, requirement of duct if any and other utilities required.
- (9) The offer must furnish preventive maintenance schedules for the liquefier and its components inclusive of mechanical, electrical and electronic systems viz. cold box, compressor, pumps, meters, switch gears, control circuit etc.
- (10) The offer must contain the shape, size, weight and utility requirement of each equipment, component and instrument along with its suggested layout, piping arrangement, wiring diagram, required site conditions etc., as the case may be.
- (11) The offer must contain shipment packing size, weight and volume of each equipment and component.
- (12) The offer must list all the utilities that are expected from the Institute viz. Electric power (KVA/KW load of each equipment), Chilled water (pressure, flow rate and temperature rise i.e. tonnage of refrigeration consumed), Atmospheric cooling air (Fan specifications, Duct sizes, Free space required around units etc., if required), Type of floor, Ceiling height, Lifting arrangements for installation and disassembly for maintenance (Clarify the method of lifting and the ceiling height needed at the time of installation and for maintenance at a later stage).
- (13) The offer must confirm that the warrantee will start from the date of satisfactory commissioning in the Institute.
- (14) The offer must state delivery time, payment terms, warranty terms, AMC terms, number of years that the spares support is likely to be available, Type of after-sale and maintenance support available from the suppliers and their representatives in India.
- (15) The supplier should have supplied at least one similar plants in India with the necessary documents should be supported with the offer. Offers not complying this clause will not be considered.

- (16) The offer must contain the names and addresses of the organizations where similar liquefiers have been supplied in recent years in India and abroad along with the respective commissioning dates and contact details of the concerned person/so
- (17) Validity of the offer should be 120 days from the date of opening of the tender

OPTIONAL ITEMS - TO BE QUOTED SEPARATELY.

(1) Mass Spectrometer Leak Detector with moving trolley and handle

Tender Specifications

a) Smallest detectable helium leak [Vacuum mode] : ~5x10-12 mbar. I/Sec

b) Smallest detectable helium leak [Sniffer mode] : ~5x10-8 mbar. I/Sec

c) Maximum detectable helium leak rate : ~1.0 mbar. I/Sec

d) Detectable Masses : Hydrogen, Helium-3

& Helium -4

- e) External calibrated rate of about 10-8 mbar. I/Sec
- f) Sniffer line of about 5-meter length with sniffer handle and probe.
- g) Helium gas pistol with flexible hose piping.
- h) Spare Filament for the above spectrometer
- i) Overall seal Kit for fore vacuum pump.
- j) Quick coupling for vent connection.
- k) Maintenance Kit (5000 hours) for the offered model

(2) Supply of SIX (6) Nos of Liquid Helium Transport Dewars:

Tender Specifications: Vertical type, 100 liters capacity, with all accessories including pressure gauges, shutoff valves, safety valves, moving handles, roller base with clamping castors, pump-out port and ready to use.

(3) Supply of SIX (6) Nos. of Liquid Nitrogen Transport Dewars:

Tender Specifications: ~ 250 liters capacity, vertical type, non-magnetic, all Stainless Steel, Square base with pull handle with roller castors, level indicator, ready to use with all necessary safety devices integrated in the dewar.

III. ATOMIC FORCE MICROSCOPE SCANNING TUNNELING MICROSCOPE - 1 NO SCHEDULE

1. TENDER NO. : IISER/PUR/5659/14

2. Due Date : 26th May 2015 [4 PM]

3. Date of Opening of : 27th May 2015 [3 PM]

TECHNICAL BID

4. DATE OF OPENING OF PRICE : Will be notified later.

BID

5. Cost of Tender # : Rs. 500 + 4.04% VAT

6. EMD [BID SECURITY] ** : Rs. 2 Lakhs

TECHNICAL SPECIFICATIONS FOR AFM/STM

AFM Contact mode both in air and liquid				
Non-Contact AFM mode/Intermittent contact mode/tapping mode in				
both in air and liquid				
Lateral Force Microscopy				
Phase Imaging				
Software or hardware solutions for studying delicate soft samples and				
imaging under liquid environments without compromising resolution				
and clarity				
Force spectroscopy				
Point and shoot force mapping (1nm resolution or less)				
Magnetic force microscopy				
Piezo Response Force Microscopy				
Electrostatic force microscopy				
High-speed methodology (hardware or software) for studying				
molecular binding events and simultaneous Topography and				
Adhesion data				
Kelvin Probe Microscopy				
Amplitude and frequency modulation				
Conductive AFM				
Capability to perform conductive measurements while scanning as well as at user specified locations (I-V curves)				
The software must allow user-specified wave forms for I/V				
spectroscopy (square, sine, triangle, pulse, or user defined)				
The current sensing ability as low as 1 pA or better and should be compatible for soft samples without any sample damage				
STM attachment should fit directly onto existing scanner assembly				
and should be easy to operate with software selection				
atomic resolution imaging capability in the ambient (to be				
demonstrated)				
constant current and constant height feedback modes				
tip sample approach / engagement at small currents ~ 100 pA or lower				
capable of stable imaging with tunnel current setpoint in the range 100 pA – 10 nA or better				
user selectable sample/tip bias in the range few mV to ±10 V with smallest step size preferably ~ 5mV				

	T
	selectable I-V convertor amplification factor
	Ability to perform measurements under air and liquid environments
	user selectable feedback modes (e.g. linear, logarithmic, etc.)
Scanner	XY scan range minimum 90 μM x 90 μM or above
	Z scan range minimum 5 µM with noise ≤ 0.03 nm or lower
	XY scan range 8 μM x 8 μM or above (upto 15 μM x 15 μM)
	Z scan range 2 μ M or above with noise level ≤ 0.03 nm or lower
	XY scan range 1 μM x 1 μM or below Z scan range 0.3 μM or higher with noise level ≤ 0.006 nm or lower
	The same scanner should be able to operate under the specified
	temperature range, ambient or lower to +250 °C, in all modes without
	compromising specified scan range and noise values
	The scanner should have protection against liquids and should be
	compatible with various solvents including water and hydrophobic
	solvents (dichlorobenzene etc.)
	Manimum Image pixel resolution: 4096 x 4096 or above
	Maimian mage pixer resolution. 1030 x 1030 of above
Sample Stage	Typical sample size 10 mm x 10 mm x 5 mm or larger
	Accessories for controlling the sample temperature ambient or lower
	to +100 °C (or above) with stability minimum 0.1 °C
	Good temp stability
	Accessories for imaging under liquid environments
	Accessories for environmental control
	Microscope optics should have 1 Megapixel or better camera and
	required accessories for visualizing sample surface, tip etc. The
	camera should be connected to a computer with software for
	capturing the images
Controller	Latest high speed controller with low noise having access to various input and out signals to/from the SPM controller
	High-frequency scanner for fast scanning with higher pixel resolution
	User selectable image resolution up to 4096 x 4096 or better
	Tuning of the cantilever resonance up to 2 MHz
	Simultaneous display and capture of up to eight images
	Simultaneous display and capture of up to cigit images
	RNC option for various input and output signals to from SPM
	BNC option for various input and output signals to/from SPM
	controller
	controller 16 bit or higher digital signal processor
	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs
	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs
	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs
	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs
	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements
	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control
Light source	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc.
Light source	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source
Light source and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength
_	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30'
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength
and detector	Controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30' LED monitors Preferably Intel core i7 quad processor or equivalent processor
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30' LED monitors Preferably Intel core i7 quad processor or equivalent processor
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30' LED monitors Preferably Intel core i7 quad processor or equivalent processor 4 GB RAM or higher Internal hard drive space of 500 GB or higher capacity in two
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30' LED monitors Preferably Intel core i7 quad processor or equivalent processor 4 GB RAM or higher Internal hard drive space of 500 GB or higher capacity in two partitions
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30' LED monitors Preferably Intel core i7 quad processor or equivalent processor 4 GB RAM or higher Internal hard drive space of 500 GB or higher capacity in two partitions operating system win 7 or higher/LINUX
and detector	controller 16 bit or higher digital signal processor Nine or more 16 bit or higher ADCs Three or more 16 bit or higher DACs Three or more lock-in amplifiers for simultaneous measurements Option for open and closed loop scanning Digital Q control Realtime adjustable scan rate, size, offset, bias voltage, current etc. Low coherence laser light source position sensitive quadreplexed photodetector optimized for laser light wavelength Latest computer with Windows/LINUX based operating system ad 30' LED monitors Preferably Intel core i7 quad processor or equivalent processor 4 GB RAM or higher Internal hard drive space of 500 GB or higher capacity in two partitions

Software -Windows/LINUX based software for controlling the SPM, image acquisition and processing Image processing software must have the common image processing features like surface roughness, cross section, grain size, depth, power spectral, histogram, fractal, fourier, image mixing, auto correlation, enhanced image filtering, force distance curve, 2D fast fourier, plane fit, high pass and low pass, zoom in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Active air -filled anti-vibration table with pneumatic assembly must be provided		Complete existem (windows, and other coftware) distra
acquisition and processing Image processing software must have the common image processing features like surface roughness, cross section, grain size, depth, power spectral, histogram, fractal, fourier, image mixing, auto correlation, enhanced image filtering, force distance curve, 2D fast fourier, plane fit, high pass and low pass, zoom in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling	Coftmag	Complete system (windows, spm and other software) disks
Image processing software must have the common image processing features like surface roughness, cross section, grain size, depth, power spectral, histogram, fractal, fourier, image mixing, auto correlation, enhanced image filtering, force distance curve, 2D fast foureir, plane fit, high pass and low pass, zoom in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items The cost three levels of training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling	Sonware	,
features like surface roughness, cross section, grain size, depth, power spectral, histogram, fractal, fourier, image mixing, auto correlation, enhanced image filtering, force distance curve, 2D fast foureir, plane fit, high pass and low pass, zoom in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsic installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 10 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
power spectral, histogram, fractal, fourier, image mixing, auto correlation, enhanced image filtering, force distance curve, 2D fast foureir, plane fit, high pass and low pass, zoom in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items Other items The of cost three levels of training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 1 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
correlation, enhanced image filtering, force distance curve, 2D fast foureir, plane fit, high pass and low pass, 200m in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Active air -filled anti-vibration table with pneumatic assembly must be provided Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
foureir, plane fit, high pass and low pass, zoom in/out, variable shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes- 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
shading and display angle tilt, 2D and 3D height representation. Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereo Microscope Microscope Stereo Microscope Microscope Anti-vibration Other items Other items Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 1 µM Calibration grid 1 µM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Additional labview drivers for custom programming and controlling the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
the instrument Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Possibility for Scripting data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereon Stereo and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Calibration grid 10 µM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
data export to ascii, bmp, jpg, tiff, matlab etc. Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items Other items Tree of cost three levels of training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 11 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Must include free AFM software upgrades for the life of the instrument Stereo Microscope Stereomicroscoepe with 2.8 MP or better microscope digital camera and minimum of 100X magnification for viewing the AFM tip and sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		Possibility for Scripting
Stereo Microscope Microscope Microscope Microscope Microscope Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		data export to ascii, bmp, jpg, tiff, matlab etc.
Stereo Microscope Microscope Microscope Microscope Microscope Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 1 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		Must include free AFM software upgrades for the life of the
Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Other items Other items Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		instrument
Sample surface Anti-vibration Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 11 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling	Stereo	Stereomicroscoepe with 2.8 MP or better microscope digital camera
Active air -filled anti-vibration table with pneumatic assembly must be provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling	Microscope	and minimum of 100X magnification for viewing the AFM tip and
Deb provided Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 10 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		sample surface
Other items Onsite installation and training of equipment and demonstration of all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling	Anti-vibration	Active air -filled anti-vibration table with pneumatic assembly must
all modes purchased on typical samples Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		be provided
Free of cost three levels of training of duration atleast one week each on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes - 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling	Other items	Onsite installation and training of equipment and demonstration of
on (i) basic modes at the time of installation (ii) basic + advanced level training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes - 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 10 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		all modes purchased on typical samples
training after four months and (iii) basic + advanced level training after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		Free of cost three levels of training of duration atleast one week each
after one year Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 1 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		on (i) basic modes at the time of installation (ii) basic + advanced level
Demonstration of XYZ calibration with calibration samples for AFM and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		training after four months and (iii) basic + advanced level training
and STM Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 1 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		after one year
Contact mode Probes – 10 No's Tapping Probes- 100 No's Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		_
Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 1 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Tapping Probes (high resolution tip) - 100 No's Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 1 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Conductive mode Probes – 50 No's Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Kelvin Force Probes – 50 No's STM probe-20 Nos Calibration grid 10 μM Calibration grid 1 μM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
STM probe-20 Nos Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Calibration grid 10 µM Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Calibration grid 1 µM Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Standard for Kelvin Probe work function measurements HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
HOPG for STM Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Rest can be brought for demo All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
All equipment should be in compliance with Indian Regulations. Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
Electrical and electronic equipment should be for 230V, 50Hz AC supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
supply. The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
The Indian agents should have service facilities in India and should have factory trained and qualified engineers to attend on calling		
have factory trained and qualified engineers to attend on calling		
Undertaking of service and Support		
		Undertaking of service and Support
1 Year standard warranty		1 Year standard warranty
4 years of additional warranty		4 years of additional warranty
Vendors should undertake the availability of all the parts for at least		
10 years		▼

Vendors should provide the technical brochures to support their compliance statement

TERMS & CONDITIONS

- 1) Separate Bid with Tender No. & Item Name shall be separately sent for each Tender
- **2)** Instructions to Bidders [Annexure I] and General Terms and Conditions of the Contract [Annexure II] shall be followed.
- 3) #Tender cost in the form of DD favouring IISERTVM shall be sent along with 'Technical bid'.
- **4)** **Bid Security [refundable without any interest] shall be as per clause 1.15 of Annexure I and shall also be sent alongwith Technical bid. Incase the Bid security is provided by Bank Guarantee, specimen for the same is at Annexure V.
- 5) Integrity Pact [Annexure III] shall be completed, signed, sealed and sent alongwith Technical Bid.
- **6)** Clause 1.10[Bid Form] of Annexure I is **not** applicable for this Tender.
- 7) <u>Tenders without Tender Cost and Bid Security shall not</u> <u>be considered and summarily to be disqualified from participating tender.</u>
- 8) Tender Opening:

All tenders will be opened at GEM Building. Authorized representatives [with authorization letters] of the bidders may attend the Tender Opening.

Two - Part Tender

PART I – TECHNICAL BID:

Consisting of all Technical details alongwith commercial terms and conditions. DD towards Tender Cost, Bid Security [EMD] and Integrity Pact signed shall be sent alongwith Technical bids, Clientele list and details of TWO major installations in India.

PART II - PRICE BID:

Showing item-wise prices for the items mentioned in the Technical Bid.

The technical bid and the Price bid should be in separate sealed and superscribed covers. Both these covers are to be put in a sealed bigger cover superscribing Tender No and due date. Technical bids will be opened first and after evaluation, price bids of only the shortlisted bidders will be opened.

Sd/DEPUTY REGISTRAR
[PURCHASE & STORES]
IISERTVM

INSTRUCTIONS TO BIDDERS

1.1. ELIGIBLE BIDDERS.

- 1.1.1 This Invitation for Bids is open to all suppliers.
- 1.1.2 Bidders should not be associated, or have been associated in the past, directly or indirectly, with a firm or any of its affiliates which have been engaged by the Purchaser to provide consulting services for the preparation of the design, specifications, and other documents to be used for the procurement of the goods to be purchased under this Invitation of Bids.

1.2. Cost of Bidding

1.2.1 The Bidder shall bear all costs associated with the preparation and submission of its bid, and "the Purchaser", will in no case be responsible or liable for these costs, regardless of the conduct or outcome of the bidding process.

1.3. FRAUD AND CORRUPTION:

- 1.3.1 The purchaser requires that the bidders/suppliers and contractors observe the highest standard of ethics during the procurement and execution of such contracts.
- 1.3.2 The purchaser will reject a proposal for award if it determines that the Bidder recommended for award has, directly or through an agent, engaged in corrupt, fraudulent, collusive or coercive practices in competing for the Contract in question;

 The Bidding Documents

1.4. Cost of Bidding Documents

1.4.1 Interested eligible bidders may download the document from our Website. However the bid document cost to be paid during the submission of the bid.

1.5. CONTENT OF BIDDING DOCUMENTS

- 1.5.1 The goods required, bidding procedures and contract terms are prescribed in the bidding documents which should be read in conjunction.
- 1.5.2 The Bidder is expected to examine all instructions, forms, terms, and specifications in the bidding documents. Failure to furnish all information required by the bidding documents or submission of a bid not substantially responsive to the bidding documents in every respect will be at the Bidder's risk and may result in rejection of its bid.

1.6. CLARIFICATION OF BIDDING DOCUMENTS

A prospective Bidder requiring any clarification of the Bidding Documents shall contact the Purchaser in writing via fax/email. The Purchaser will respond in writing to any request for clarification, provided that such request is received not later than ten (10) days prior to the deadline for submission of bids. The Purchaser shall forward copies of its response to all those who have acquired the Bidding Documents directly from it, including a description of the inquiry but without identifying its source. Should the Purchaser deem it necessary to amend the Bidding Documents as a result of a clarification, it shall do so following the procedure under clause relating to amendment of Bidding documents and Clause relating to Deadline for Submission of Bids. The clarifications and amendments issued would also be hosted on the website of the purchaser for the benefit of the other prospective bidders.

1.7. AMENDMENT OF BIDDING DOCUMENTS

- 1.7.1 At any time prior to the deadline for submission of bids, the Purchaser may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder, modify the bidding documents by amendment.
- 1.7.2 All prospective bidders who have received the bidding documents will be

- notified of the amendment in writing or by e-mail and will be binding on them. The same would also be hosted on the website of the purchaser and all prospective bidders are expected to surf the website before submitting their bids to take cognizance of the amendments.
- 1.7.3 In order to allow prospective bidders reasonable time to take the amendment into account, while in preparing their bids, the Purchaser, at its discretion, may extend the deadline for the submission of bids and host the changes on the website of the purchaser.
- 1.7.4 In case of any amendment to the bid, it will be done at least 7 days before due date and will be published in website. Bidders are required to keep this in view before submission. Normally extension of due date will not be entertained. However in special cases if due date is extended, it will be notified under the same tender notice before 3 days of due date.

PREPARATION OF BIDS

1.8. LANGUAGE OF BID

1.8.1 The bid prepared by the Bidder, as well as all correspondence and documents relating to the bid exchanged by the Bidder and the Purchaser, shall be written in English language only especially when the details are technical.

However of GOI makes it mandatory under Rajbhasha Abhniyam in that case views of Rajbasha unit of IISER-TVM may be sought.

1.8.2 The Supplier shall bear all costs of translation, if any, to the English language and all risks of the accuracy of such translation, for documents provided by the Supplier.

1.9. DOCUMENTS COMPRISING THE BID

- 1.9.1 The bid prepared by the Bidder shall include:
 - a) Bidder Information
 - b) Bid security as specified in the Invitation to Bids.
 - c) Service support details
 - d) Deviation Statement if any.
 - e) Performance Statement.
 - f) Manufacturer's Authorization.
 - g) Documentary evidence establishing that the bidder is eligible to bid and is qualified to perform the contract if its bid is accepted.
 - h) Bid form.
 - i) Documents establishing goods eligibility and conformity to bidding documents.
 - i) Applicable Price Schedule Form.
 - k) DGS&D Registration certificate in case the items under procurement falls under the restricted category of the current export-import policy of the Govt. of India.

1.10. BID FORM AND PRICE SCHEDULE

1.10.1 The bidder shall complete the Bid with the appropriate price schedule.

1.11. BID PRICES

- 1.11.1 The Bidder shall indicate on the appropriate price schedule, the unit prices and total bid prices of the goods it proposes to supply under the contract.
- Prices indicated on the price-schedule form shall be entered separately in the following manner:

(A) FOR GOODS MANUFACTURED WITHIN INDIA

- (i) The price of the goods quoted Ex-works including taxes already paid.
- (ii) VAT and other taxes like excise duty etc **[shall be shown extra]** which will be payable on the goods if the contract is awarded.

- a. The charges for inland transportation, insurance and other local services required for delivering the goods at the desired destination as specified in the price schedule.
- b. The installation, commissioning and training charges including any incidental services, if any.

(B) FOR GOODS MANUFACTURED ABROAD

Bidders are required to show break-up of Ex-works [which should include packing and handling charges], FOB/FCA and CIF/CIP price in the quotation. IISERTVM has a mechanism for consolidation and customs clearance. The consolidators will be nominated for this purpose at the time of awarding the contract if desires so.

- (i) The price of the goods, quoted on FCA (named place of delivery abroad) or FOB (named port of shipment), as specified in the price schedule.
- (ii) The charges for insurance and transportation of the goods to the port/place of destination.

(iii) The agency commission charges should be shown, if any.

- (iv) The installation, commissioning and training charges including any incidental services, if any
- 1.11.3 The terms FOB, FCA, CIF, CIP etc shall be governed by the rules prescribed in the current edition of the Incoterms published by the International Chambers of Commerce, Paris.
- 1.11.4 Where there is no mention of packing, forwarding, freight, insurance changes, taxes etc. such offer shall be rejected as incomplete.
- 1.11.5 The price quoted shall remain fixed during the contract period and shall not vary on any account
- 1.11.6 All lots and items must be listed and priced separately in the Price Schedules. If a Price Schedule shows items listed but not priced their prices shall be assumed to be included in the prices of other items. Lots or items not listed in the Price Schedule shall be assumed to be not included in the bid.
- 1.11.7 The purchases made by the purchaser for scientific purpose are exempt from excise duty under 10/97 notification and Custom Duty under notification 51/96.

1.12. BID CURRENCIES

1.12.1 Prices shall be quoted in Indian Rupees for offers received for supply within India and in freely convertible foreign currency in case of offers received for supply from foreign countries.

1.13. DOCUMENTS ESTABLISHING BIDDER'S ELIGIBILITY AND QUALIFICATIONS

- 1.13.1 The bidder shall furnish, as part of its bid, documents establishing the bidders' eligibility to bid and its qualification to perform the contract if its bid is accepted.
- 1.13.2 The documentary evidence of the bidders qualification to perform the contract if the bid is accepted shall establish to the purchasers satisfaction that;
 - (a) The bidder meets the qualification criteria listed in bidding documents, if any.
 - (b) Bidder that doesn't manufacture the goods it offers to supply shall submit to Manufacturers' Authorization Form (MAF) to demonstrate that it has been duly authorized by the manufacturer of the goods to quote and/or supply the goods.
 - (c) In case a bidder not doing business within India, it shall furnish the certificate to the effect that the bidder is or will be represented by an agent in India equipped and able to carry out the supply, maintenance, repair obligations etc. during the warranty and post warranty period or ensure a mechanism at place for carrying out the supply, maintenance, repair

- obligations etc. during the warranty and post-warranty period.
- 1.13.3 Conditional tenders shall not be accepted

1.14. DOCUMENTS ESTABLISHING GOODS' ELIGIBILITY AND CONFORMITY TO BIDDING DOCUMENTS

- 1.14.1 To establish the goods' eligibility, the documentary evidence of the goods and services eligibility shall consist of a statement on the country of origin of the goods and services offered which shall be confirmed by a certificate of origin at the time of shipment.
- 1.14.2 To establish the conformity of the goods and services to the specifications and schedule of requirements of the bidding document, the documentary evidence of conformity of the goods and services to the bidding documents may be in the form of literature, drawings and data, and shall consist of:
 - (a) A detailed description of the essential technical and performance characteristics of the goods;
 - (b) A list giving full particulars, including available sources and current prices, of spare parts, special tools, etc., necessary for the proper and continuing functioning of the goods during the warranty period following commencement of the use of the goods by the Purchaser in the Priced bid; and
 - (c) An item-by-item commentary on the Purchaser's Technical Specifications demonstrating substantial responsiveness of the goods and services to those specifications or a statement of deviations and exceptions to the provisions of the Technical Specifications.
- 1.14.3 For purposes of the commentary to be furnished pursuant to above, the Bidder shall note that standards for workmanship, material and equipment, designated by the Purchaser in its Technical Specifications are intended to be descriptive only and not restrictive. The Bidder may substitute these in its bid, provided that it demonstrates to the Purchaser's satisfaction that the substitutions ensure substantial equivalence to those designated in the Technical Specifications.

1.15. BID SECURITY

- 1.15.1 The Bidder shall furnish, as part of its bid, a bid security (BS) for an amount as specified in the Invitation for Bids. In the case of foreign bidders, the BS shall be submitted either by the principal or by the Indian agent and in the case of indigenous bidders; the BS shall be submitted by the manufacturer or their specifically authorized dealer/bidder.
- 1.15.2 The bid security shall be in Indian Rupees for offers received for supply within India and denominated in the currency of the bid or in any freely convertible foreign exchange in the case of offers received for supplies from foreign countries in equivalent Indian Rupees. The bid security shall be in one of the following forms at the bidders' option:
 - (a) A bank guarantee issued by a Nationalized/Scheduled bank/Foreign Bank operating in India in the form provided in the bidding documents and valid for 45 days beyond the validity of the bid. In case a bidder desires to submit a BG issued from a foreign bank, then the same should be confirmed by a Nationalized/Scheduled Indian bank; or
 - (b) Fixed Deposit receipt pledged in favour of the IISER-TVM.
 - (c) A Banker's cheque or demand draft in favour of IISERTVM issued by any Nationalised/Scheduled Indian bank.
- 1.15.3 The bid security shall be payable promptly upon written demand by the purchaser in case the conditions listed in the ITB clause 15.11 are invoked.

- 1.15.4 The bid security should be submitted in its original form. Copies shall not be accepted.
- 1.15.5 While Bid security (EMD) is a requirement, the Director IISER-TVM may grant exemption of Bid security to some specific parties having sound credentials and are of national/international repute.
- 1.15.6 The bid security of unsuccessful bidder will be discharged /returned as promptly as possible positively within a period of 15 days after the expiration of the period of bid validity or placement of order whichever is later, without any interest.
- 1.15.7 The successful Bidder's bid security will be discharged upon the Bidder furnishing the performance security, without any interest. Alternatively, the BS could also be adjusted against PS, if it is paid through DD/BC.
- 1.15.8 The firms registered with DGS&D, NSIC if any, are exempted from payment of bid security (BS) provided such registration includes the item they are offering which are manufactured by them and not for selling products manufactured by other companies.
- 1.15.9 In case a bidder intimates at the time of tender opening in writing that the bid security is kept inside the financial bid, then in such cases, the technical bid of the party would be accepted provisionally till opening of the financial bids with which the party has attached the bid security.
- 1.15.10 The bid security may be forfeited:
 - (a) If a Bidder withdraws or amends or impairs or derogates its bid during the period of bid validity specified by the Bidder on the Bid Form; or
 - (b) In case of a successful Bidder, if the Bidder fails to furnish order acceptance within 15 days of the order or fails to sign the contract and/or fails to furnish Performance Security within 21 days from the date of contract/ order.
- 1.16. Period of Validity of Bids
- 1.16.1 Bids shall remain valid for minimum of 90 days after the date of bid opening prescribed by the Purchaser. A bid valid for a shorter period shall be rejected by the Purchaser as non-responsive.
- In exceptional circumstances, the Purchaser may solicit the Bidder's consent to an extension of the period of validity. The request and the responses thereto shall be made in writing (or by cable, telex, fax or e-mail). The bid security provided shall also be suitably extended. A Bidder may refuse the request without forfeiting its bid security. A Bidder granting the request will not be required nor permitted to modify its bid.
- 1.16.3 Bid evaluation will be based on the bid prices without taking into consideration the above corrections.

1.17. FORMAT AND SIGNING OF BID

- 1.17.1 The bids may be submitted in single envelop or in two parts as specified in the Invitation for Bids.
- 1.17.2 In case the bids are invited on single envelop basis, then the Bidder shall prepare two copies of the bid, clearly marking each "Original Bid" and "Copy Bid", as appropriate. In the event of any discrepancy between them, the original shall govern.
- 1.17.3 In case the bids are invited on two-bid system, the Bidder shall submit the bids in two separate parts. One part shall contain Technical bid comprising all documents listed under clause relating to Documents Comprising the Bid excepting bid form and price schedules. The other part shall contain the priced-bid comprising bid form and price schedules. The Bidder shall prepare two copies of the bid, clearly marking each "Original Bid" and

- "Copy Bid", as appropriate.
- 1.17.4 The original and all copies of the bid shall be typed or written in indelible ink and shall be signed by the Bidder or a person or persons duly authorized to bind the Bidder to the Contract. All pages of the bid, except for un-amended printed literature, shall be initialed by the person or persons signing the bid.
- 1.17.5 Any interlineations, erasures or overwriting shall be valid only if they are initialed by the persons or persons signing the bid.

1.18. SUBMISSION, SEALING AND MARKING OF BIDS

- 1.18.1 The bidders may submit their duly sealed bids generally by post or by hand.
- 1.18.2 In the case of bids invited on single envelop basis, the Bidders shall seal the original and each copy of the bid in separate inner envelopes, duly marking the envelopes as "original" and "copy". The envelopes shall then be sealed in an outer envelope.
- 1.18.3 In the case of bids invited on two part basis, the Bidder shall seal the unpriced commercial and technical bid comprising the documents as listed in ITB 1.9.1 excepting for h & j and the priced bid in two separate envelops duly marked as "Technical bid" and "priced bid". Both the envelopes shall then be sealed in one outer envelope.
- 1.18.4 (a) The inner and outer envelopes shall be addressed to the Purchaser indicated.
 - (b) Bear the name and address of the bidder, Tender No., due date and a warning "Do not open before ______" to be completed with the time and date as specified in the invitation for bids.
- 1.18.5 If the outer envelope is not sealed and marked as required above, the Purchaser will assume no responsibility for the bid's misplacement or premature opening. In such cases, bids received in open condition within the due date and time will be accepted at the risk of the bidder if the same is presented to the Stores & Purchase Officer before expiry of the due date and time of opening of the bids.
- 1.18.6 Firms submitting bids in a single envelope against the requirement of two-bid system would be considered for further evaluation at the risk & responsibility of the bidder. However, the opened priced bid would be sealed immediately by the TOC without disclosing the price.

1.19. DEADLINE FOR SUBMISSION OF BIDS

- 1.19.1 Bids must be received by the Purchaser at the address specified in Invitation for bids not later than the time and date specified therein. In the event of the specified date for the submission of Bids being declared a holiday for the Purchaser, the Bids will be received up to the appointed time on the next working day.
- 1.19.2 The Purchaser may, at its discretion, extend the deadline for submission of bids by amending the bid documents in accordance with Clause relating to Amendment of Bidding Documents in which case all rights and obligations of the Purchaser and Bidders previously subject to the deadline will thereafter be subject to the deadline as extended.

1.20. LATE BIDS

- 1.20.1 Any bid received by the Purchaser after the deadline for submission of bids prescribed by the Purchaser will be rejected.
- 1.20.2 Such tenders shall be marked as late and not considered for further evaluation. They shall **not** be opened at all and be returned to the bidders in their original envelope without opening.

1.21. WITHDRAWAL, SUBSTITUTION AND MODIFICATION OF BIDS

1.21.1 A Bidder may withdraw, substitute, or modify its Bid after it has been submitted by sending a written notice in accordance with ITB Clause 18 duly signed by an authorized representative, and shall include a copy of the authorization in accordance with ITB Sub- Clause 17.4 (except that no copies of the withdrawal notice are required). The corresponding substitution or modification of the bid must accompany the respective written notice.

All notices must be:

- (a) submitted in accordance with ITB Clauses 17 and 18 (except that withdrawal notices do not require copies), and in addition, the respective envelopes shall be clearly marked "WITHDRAWAL," "SUBSTITUTION," or "MODIFICATION;" and (b) Received by the Purchaser prior to the deadline prescribed for submission of bids, in accordance with ITB Clause 19.
- 1.21.2 Bids requested to be withdrawn in accordance with ITB Sub-Clause 21.1 shall be returned unopened to the Bidders. No bid may be withdrawn, substituted, or modified in the interval between the deadline for submission of bids and the expiration of the period of bid validity specified by the Bidder on the Bid Form or any extension thereof.

 Opening and Evaluation of Bids

1.22. OPENING OF BIDS BY THE PURCHASER

- 1.22.1 The Purchaser will open all bids one at a time in the presence of Bidders' authorized representatives who choose to attend, as per the schedule given in invitation for bids. The Bidders' representatives who are present shall sign the quotation opening sheet evidencing their attendance. In the event of the specified date of Bid opening being declared a holiday for the Purchaser, the Bids shall be opened at the appointed time and location on the next working day. In two-part bidding, the financial bid shall be opened technical evaluation. 1.22.2 First, envelopes "WITHDRAWAL" shall be opened and read out and the envelope with the corresponding bid shall not be opened, but returned to the Bidder. No bid withdrawal shall be permitted unless the corresponding withdrawal notice contains a valid authorization to request the withdrawal and is read out at bid opening. Next, envelopes marked "SUBSTITUTION" shall be opened and read out and exchanged with the corresponding Bid being substituted, and the substituted Bid shall not be opened, but returned to the Bidder. No Bid substitution shall be permitted unless the corresponding substitution notice contains a valid authorization to request the substitution and is read out at bid opening. Envelopes marked "MODIFICATION" shall be opened and read out with the corresponding Bid. No Bid modification shall be permitted unless the corresponding modification notice contains a valid authorization to request the modification and is read out at Bid opening. Only envelopes that are opened and read out at Bid opening shall be considered further.
- 1.22.2 The bidders' names, bid modifications or withdrawals, bid prices, discounts, and the presence or absence of requisite bid security and such other details as the Purchaser, at its discretion, may consider appropriate, will be announced at the opening. No bid shall be rejected at bid opening, except for late bid(s). The contents of the bid forms and price schedules would however be announced only at the time of opening of Price-bids in the case of two-bid system.
- 1.22.2.1 Bids that are received late shall not be considered further for evaluation, irrespective of the circumstances.

1.22.2.2 Bidders interested in participating in the bid opening process, should depute their representatives along with an <u>authority</u> <u>letter</u> to be submitted to the purchaser at the time of bid opening.

1.23. CONFIDENTIALITY

- 1.23.1 Information relating to the examination, evaluation, comparison, and post qualification of bids, and recommendation of contract award, shall not be disclosed to bidders or any other persons not officially concerned with such process until publication of the Contract Award.
- 1.23.2 Any effort by a Bidder to influence the Purchaser in the examination, evaluation, comparison, and post qualification of the bids or contract award decisions may result in the rejection of its Bid.

1.24. CLARIFICATION OF BIDS

1.24.1 To assist in the examination, evaluation, comparison and post qualification of the bids, the Purchaser may, at its discretion, ask the Bidder for a clarification of its bid. The request for clarification and the response shall be in writing and no change in prices or substance of the bid shall be sought, offered or permitted. However, no negotiation shall be held except with the lowest bidder, at the discretion of the purchaser. Any clarification submitted by a bidder in respect to its bid which is not in response to a request by the purchaser shall not be considered.

1.25. Conversion to Single Currency

1.25.1 To facilitate evaluation and comparison, the Purchaser will convert all bid prices expressed in the amounts in various currencies in which the bid prices are payable to Indian Rupees at the selling exchange rate established by any bank in India on the date of bid opening in the case of single part bidding and the rates prevalent on the date of opening of the Priced bids in the case of two-part bidding. For this purpose, exchange rate notified in www.xe.com or www.rbi.org or any other website could also be used by the purchaser.

1.26. CONTACTING THE PURCHASER

- 1.26.1 Subject to ITB Clause 1.24, no Bidder shall contact the Purchaser on any matter relating to its bid, from the time of the bid opening to the time the Contract is awarded.
- 1.26.2 Any effort by a Bidder to influence the Purchaser in its decisions on bid evaluation, bid comparison or contract award may result in rejection of the Bidder's bid.

1.27. Post qualification

- 1.27.1 In the absence of pre-qualification, the Purchaser will determine to its satisfaction whether the Bidder that is selected as having submitted the lowest evaluated responsive bid is qualified to perform the contract satisfactorily, in accordance with the criteria listed in ITB Clause 13.
- 1.27.2 The determination will take into account the eligibility criteria listed in the bidding documents and will be based upon an examination of the documentary evidence of the Bidder's qualifications submitted by the Bidder, as well as such other information as the Purchaser deems necessary and appropriate.
- 1.27.3 An affirmative determination will be a prerequisite for award of the contract to the Bidder.

A negative determination will result in rejection of the Bidder's bid.

1.28. PURCHASER'S RIGHT TO ACCEPT ANY BID AND TO REJECT ANY OR ALL BIDS

The Purchaser reserves the right to accept or reject any bid, and to annul the bidding process and reject all bids at any time prior to award of Contract, without thereby incurring any liability to the affected Bidder or Bidders.

GENERAL TERMS & CONDITIONS

1.29. WARRANTY

- 1.29.1 The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.
- 1.29.2 The Supplier further warrants that the Goods shall be free from defects arising from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in India.
- 1.29.3 Unless otherwise specified, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated, or for eighteen (18) months after the date of shipment from the port or place of loading in the country of origin, whichever period concludes earlier.
- 1.29.4 The Purchaser shall give notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof.

 The Purchaser shall afford all reasonable opportunity for the Supplier to
 - inspect such defects.
- 1.29.5 Upon receipt of such notice, the Supplier shall, within a reasonable period of time, expeditiously repair or replace the defective Goods or parts thereof, at no cost to the Purchaser.
- 1.29.6 If having been notified, the Supplier fails to remedy the defect within the reasonable period of time; the Purchaser may proceed to take within a reasonable period such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Purchaser may have against the Supplier under the Contract.
- 1.29.7 Goods requiring warranty replacements must be replaced on free of cost basis to the purchaser.

1.29.8 Performance Guarantee:

The successful bidder is required to submit a performance Guarantee of 10% of Purchase Order value valid during warranty period.

1.30 PENALTY CLAUSE

1.30.1 Subject to GCC Clause on Force Majeure, if the Supplier fails to deliver any or all of the Goods or to perform the Services within the period(s) specified in the Contract, the Purchaser shall, without prejudice to its other remedies under the Contract, deduct from the Contract Price, as penalty, a sum equivalent to the percentage specified in contract of the delivered price of the delayed Goods or unperformed Services or contract value for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the Percentage specified in contract. Once the maximum is reached, the Purchaser may consider termination of the Contract pursuant to GCC. Clause on Termination for Default. The SCC shall also indicate the basis for ascertaining the value on which the penalty shall be applicable

1.31 INSURANCE

- 1.31.1 Should the purchaser elect to buy on CIF/CIP basis, the Goods supplied under the Contract shall be fully insured in Indian Rupees against any loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the manner specified in contract.
- 1.31.2 Where delivery of the goods is required by the purchaser on CIF or CIP basis the supplier shall arrange and pay for Cargo Insurance, naming the

- purchaser as beneficiary and initiate & pursue claims till settlement, on the event of any loss or damage.
- 1.31.3 Where delivery is on FOB or FCA basis, insurance would be the responsibility of the purchaser.
- 1.31.4 With a view to ensure that claims on insurance companies, if any, are lodged in time, the bidders and /or the Indian agent shall be responsible for follow up with their principals for ascertaining the dispatch details and informing the same to the Purchaser and he shall also liaise with the Purchaser to ascertain the arrival of the consignment after clearance so that immediately thereafter in his presence the consignment could be opened and the insurance claim be lodged, if required, without any loss of time. Any delay on the part of the bidder/Indian Agent would be viewed seriously and he shall be directly responsible for any loss sustained by the purchaser on the event of the delay.

1.32 TRANSPORTATION

- 1.32.1 Where the Supplier is required under the Contract to deliver the Goods FOB, transport of the Goods, up to and including the point of putting the Goods on board the vessel at the specified port of loading, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract price. Where the Supplier is required under the Contract to deliver the Goods FCA, transport of the Goods and delivery into the custody of the carrier at the place named by the Purchaser or other agreed point shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract price.
- 1.32.2 Where the Supplier is required under the Contract to deliver the Goods CIF or CIP, transport of the Goods to the port of destination or such other named place of destination in the Purchaser's country, as shall be specified in the Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.
- 1.32.3 In the case of supplies from within India, where the Supplier is required under the Contract to transport the Goods to a specified destination in India, defined as the Final Destination, transport to such destination, including insurance and storage, as specified in the Contract, shall be arranged by the Supplier, and the related costs shall be included in the Contract Price.

1.33 PACKING

Should be as per standards prescribed for international trade and movement of the goods.

1.34 **STANDARDS**

The Goods supplied and services rendered under this Contract shall conform to the standards mentioned in the Technical Specifications and Schedule of Requirements, and, when no applicable standard is mentioned, to the authoritative standard appropriate to the Goods' country of origin and such standards shall be the latest issued by the concerned institution.

1.35 Performance Security

- 1.35.1 Within 21 days of receipt of the notification of award/PO, the Supplier shall furnish performance security of 7% of Contract/ Purchase Order value in the amount specified in the Contract, valid till 60 days beyond the completion of contractual obligations.
- 1.35.2 The proceeds of the performance security shall be payable to the Purchaser as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.

- 1.35.3 The Performance Security shall be denominated in Indian Rupees for the offers received for supplies within India and denominated in the currency of the contract in the case of offers received for supply from foreign countries.
- 1.35.4 In the case of imports, the PS may be submitted either by the principal or by the Indian agent and, in the case of purchases from indigenous sources, the PS may be submitted by either the manufacturer or their authorized dealer/bidder.
- 1.35.5 The Performance security shall be in one of the following forms:
 - (a) A Bank guarantee or stand-by Letter of Credit issued by a Nationalized/Scheduled bank located in India or a bank located abroad in the form provided in the bidding documents. Or,
 - (b) A Banker's cheque or Account Payee demand draft in favour of the purchaser. Or,
 - (c) A Fixed Deposit Receipt pledged in favour of the Purchaser.

1.36 Delivery Schedule

Delivery schedule should be specified clearly.

1.37 PAYMENT TERMS

For Indigenous supplies payment will be made within 30 days from the date of supply, installation and commissioning. For Imported goods, the preferred payment mode of the Institute is Sight Draft /Net 30 days. If Letter of Credit is insisted by the vendors, the LC will be established for 100% of Purchase Order value out of which 90% will be paid against shipping evidence and 10% after installation and commissioning and on submission of Performance Bank Guarantee for 10% of Purchase Order value which should be valid during warranty period. However the advance payment to the vendors is not encouraged.

Sd/-

Deputy Registrar [Purchase & Stores]

GENERAL TERMS & CONDITIONS OF THE CONTRACT

Table of Contents

S1.No.	Contents
2.1	Definitions
2.2	Contract Documents
2.3	Fraud and Corruption
2.4	Joint Venture, Consortium or Association
2.5	Scope of Supply
2.6	Suppliers' Responsibilities
2.7	Contract price
2.8	Copy Right
2.9	Application
2.10	Standards
2.11	Use of Contract Documents and Information
2.12	Patent Indemnity
2.13	Performance Security
2.14	Inspections and Tests
2.15	Packing
2.16	Delivery and Documents
2.17	Insurance
2.18	Transportation
2.19	Incidental Services
2.20	Spare Parts
2.21	Warranty
2.22	Terms of Payment
2.23	Change Orders and Contract Amendments
2.24	Assignment
2.25	Subcontracts
2.26	Extension of time
2.27	Penalty Clause
2.28	Termination for Default
2.29	Force Majeure
2.30	Termination for insolvency
2.31	Termination for Convenience
2.32	Settlement of Disputes
2.33	Governing Language
2.34	Applicable Law
2.35	Notice
2.36	Taxes and Duties
2.37	Right to use Defective Goods
2.38	Protection against Damage
2.39	Site preparation and installation

2.1 DEFINITIONS

- 2.1.1 The following words and expressions shall have the meanings hereby assigned to them:
 - a) "Contract" means the Contract Agreement entered into between the Purchaser and the Supplier, together with the Contract Documents referred to therein, including all attachments, appendices, and all documents incorporated by reference therein.
 - b) "Contract Documents" means the documents listed in the Contract Agreement, including any amendments thereto.
 - c) "Contract Price" means the price payable to the Supplier as specified in the Contract Agreement, subject to such additions and adjustments thereto or deductions there from, as may be made pursuant to the Contract.
 - d) "Day" means calendar day.
 - e) "Completion" means the fulfilment of the Related Services by the Supplier in accordance with the terms and conditions set forth in the Contract.
 - f) "GCC" means the General Conditions of Contract.
 - g) "Goods" means all of the commodities, raw material, machinery and equipment, and/or other materials that the Supplier is required to supply to the Purchaser under the Contract.
 - h) "Related Services" means the services incidental to the supply of the goods, such as transportation, insurance, installation, training and initial maintenance and other such obligations of the Supplier under the Contract.
 - i) "SCC" means the Special Conditions of Contract.
 - j) "Subcontractor" means any natural person, private or government entity, or a combination of the above, to whom any part of the Goods to be supplied or execution of any part of the Related Services is subcontracted by the Supplier.
 - k) "Supplier" means the natural person, private or government entity, or a combination of the above, whose bid to perform the Contract has been accepted by the Purchaser and is named as such in the Contract Agreement.
 - l) Indian Institute of Science Education and Research, Thiruvananthapuram (IISER-TVM) means a society registered under the Travancore Cochin Literary Scientific and Charitable Societies' Registration Act, 1995 (12 of 1955) on 20.02.2008 at Kerala at no. T.342/08.
 - m) "The final destination," where applicable, means the place named in the SCC.

2.2 CONTRACT DOCUMENTS

2.2.1 Subject to the order of precedence set forth in the Contract Agreement, all documents forming the Contract (and all parts thereof) are intended to be correlative, complementary, and mutually explanatory. The Contract Agreement shall be read as a whole.

2.3 Fraud and Corruption

- 2.3.1 The purchaser requires that bidders, suppliers, contractors and consultants, if any, observe the highest standard of ethics during the procurement and execution of such contracts. In pursuit of this policy,
 - a) The terms set forth below are defined as follows:
 - i) "Corrupt practice" means the offering, giving, receiving, or soliciting, directly or indirectly, of anything of value to influence the action of a public official in the procurement process or in contract execution;

- ii) "Fraudulent practice" means a misrepresentation or omission of facts in order to influence a procurement process or the execution of a contract;
- iii) "Collusive practice" means a scheme or arrangement between two or more bidders, with or without the knowledge of the Borrower, designed to establish bid prices at artificial, noncompetitive levels; and
- iv) "Coercive practice" means harming or threatening to harm, directly or indirectly, persons or their property to influence their participation in the procurement process or affect the execution of a contract;
- b) the purchaser will reject a proposal for award if it determines that the Bidder recommended for award has, directly or through an agent, engaged in corrupt, fraudulent, collusive or coercive practices in competing for the Contract in question;

2.4 Joint Venture, Consortium or Association

2.4.1 If the Supplier is a joint venture, consortium, or association, all of the parties shall be jointly and severally liable to the Purchaser for the fulfilment of the provisions of the Contract and shall designate one party to act as a leader with authority to bind the joint venture, consortium, or association. The composition or the constitution of the joint venture, consortium, or association shall not be altered without the prior consent of the Purchaser.

2.5 SCOPE OF SUPPLY

2.5.1 The Goods and Related Services to be supplied shall be as specified in the Schedule of Requirements.

2.6 Suppliers' Responsibilities

2.6.1 The Supplier shall supply all the Goods and Related Services included in the Scope of Supply in accordance with Scope of Supply Clause of the GCC, and the Delivery and Completion Schedule, as per GCC Clause relating to delivery and document.

2.7 CONTRACT PRICE

2.7.1 Prices charged by the Supplier for the Goods supplied and the Related Services performed under the Contract shall not vary from the prices quoted by the Supplier in its bid.

2.8 COPYRIGHT

2.8.1 The copyright in all drawings, documents, and other materials containing data and information furnished to the Purchaser by the Supplier herein shall remain vested in the Supplier, or, if they are furnished to the Purchaser directly or through the Supplier by any third party, including suppliers of materials, the copyright in such materials shall remain vested in such third party.

2.9 APPLICATION

2.9.1 These General Conditions shall apply to the extent that they are not superseded by provisions in other parts of the Contract.

2.10 STANDARDS

2.10.1 The Goods supplied and services rendered under this Contract shall conform to the standards mentioned in the Technical Specifications and Schedule of Requirements, and, when no applicable standard is mentioned, to the authoritative standard appropriate to the Goods' country of origin and such

standards shall be the latest issued by the concerned institution.

2.11 USE OF CONTRACT DOCUMENTS AND INFORMATION

- 2.11.1 The Supplier shall not, without the Purchaser's prior written consent, disclose the Contract, or any provision thereof, or any specification, plan, drawing, pattern, sample or information, furnished by or on behalf of the Purchaser in connection therewith, to any person other than a person employed by the Supplier in performance of the Contract.
 - Disclosure to any such employed person shall be made in confidence and shall extend only so far, as may be necessary for purposes of such performance.
- 2.11.2 The Supplier shall not, without the Purchaser's prior written consent, make use of any document or information enumerated above except for purposes of performing the Contract.
- 2.11.3 Any document, other than the Contract itself, enumerated above shall remain the property of the Purchaser and shall be returned (in all copies) to the Purchaser on completion of the Supplier's performance under the Contract if so required by the Purchaser.

2.12 PATENT INDEMNITY

- 2.12.1 The Supplier shall, subject to the Purchaser's compliance with GCC Sub-Clause 12.2, indemnify and hold harmless the Purchaser and its employees and officers from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Purchaser may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract by reason of:
 - (a) the installation of the Goods by the Supplier or the use of the Goods in India; and
 - (b) the sale in any country of the products produced by the Goods.
- 2.12.2 If any proceedings are brought or any claim is made against the Purchaser, the Purchaser shall promptly give the Supplier a notice thereof, and the Supplier may at its own expense and in the Purchaser's name conduct such proceedings or claim and any negotiations for the settlement of any such proceedings or claims.

2.13 Performance Security

- 2.13.1 Within 21 days of receipt of the notification of award/PO, the Supplier shall furnish performance security in the amount specified in SCC, valid till 60 days after the warranty period. Alternatively, the PS may also be submitted at the time of release of final payment in cases where part payment is made against delivery & part on installation. The PS, where applicable, shall be submitted in advance for orders where full payment is to be made on Letter of Credit (LC) or on delivery. In this case, submission of PS at the time of negotiation of documents through Bank would be stipulated as a condition in the LC and the BS should be kept valid till such time the PS is submitted.
- 2.13.2 The proceeds of the performance security shall be payable to the Purchaser as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.

- 2.13.3 The Performance Security shall be denominated in Indian Rupees for the offers received for supplies within India and denominated in the currency of the contract in the case of offers received for supply from foreign countries.
- 2.13.4 In the case of imports, the PS may be submitted either by the principal or by the Indian agent and, in the case of purchases from indigenous sources, the PS may be submitted by either the manufacturer or their authorized dealer/bidder.
- 2.13.5 The Performance security shall be in one of the following forms:
 - (a) A Bank guarantee or stand-by Letter of Credit issued by a Nationalized/Scheduled bank located in India or a bank located abroad in the form provided in the bidding documents. Or,
 - (b) A Banker's cheque or Account Payee demand draft in favour of the purchaser. Or,
 - (c) A Fixed Deposit Receipt pledged in favour of the Purchaser.
- 2.13.6 The performance security will be discharged by the Purchaser and returned to the Supplier not later than 60 days following the date of completion of the Supplier's performance obligations, including any warranty obligations, unless specified otherwise in SCC, without levy of any interest.
- 2.13.7 In the event of any contract amendment, the supplier shall, within 21 days of receipt of such amendment, furnish the amendment to the performance security, rendering the same valid for the duration of the contract, as amended for further period of 60 days thereafter.
- 2.13.8 The order confirmation should be received within 15 days from the date of notification of award. However, the Purchaser has the powers to extend the time frame for submission of order confirmation and submission of Performance Security (PS). Even after extension of time, if the order confirmation /PS are not received, the contract shall be cancelled and limited tenders irrespective of the value would be invited from the responding firms after forfeiting the bid security of the defaulting firm, where applicable provided there is no change in specifications. In such cases the defaulting firm would not be considered again for re-tendering in the particular case.

2.14. Inspections and Tests

- 2.14.1 The Supplier shall at its own expense and at no cost to the Purchaser carry out all such tests and/or inspections of the Goods and Related Services as are specified in the SCC or as discussed and agreed to during the course of finalization of contract.
- 2.14.2 The Purchaser or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Contract specifications at no extra cost to the Purchaser. The Technical Specifications and SCC shall specify what inspections and tests the Purchaser requires and where they are to be conducted. The Purchaser shall notify the Supplier in writing in a timely manner of the identity of any representatives retained for these purposes
- 2.14.3 The inspections and tests may be conducted on the premises of the Supplier or its subcontractor(s), at the point of delivery and/or at the Goods final destination. If conducted on the premises of the Supplier or its subcontractor(s), all reasonable facilities and assistance, including access to drawings and production data shall be furnished to the inspectors at no charge to the Purchaser.

- 2.14.4 Whenever the Supplier is ready to carry out any such test and inspection, it shall give a reasonable advance notice, including the place and time, to the Purchaser. The Supplier shall obtain from any relevant third party or manufacturer any necessary permission or consent to enable the Purchaser or its designated representative to attend the test and/or inspection
- 2.14.5 Should any inspected or tested Goods fail to conform to the specifications, the Purchaser may reject the goods and the Supplier shall either replace the rejected Goods or make alterations necessary to meet specification requirements free of cost to the Purchaser.
- 2.14.6 The Purchaser's right to inspect, test and, where necessary, reject the Goods after the Goods' arrival at final destination shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed by the Purchaser or its representative prior to the Goods shipment.
- 2.14.7 The Supplier shall provide the Purchaser with a report of the results of any such test and/or inspection.
- 2.14.8 With a view to ensure that claims on insurance companies, if any, are lodged in time, the bidders and /or the Indian agent shall be responsible for follow up with their principals for ascertaining the dispatch details and informing the same to the Purchaser and he shall also liaise with the Purchaser to ascertain the arrival of the consignment after clearance so that immediately thereafter in his presence the consignment could be opened and the insurance claim be lodged, if required, without any loss of time. Any delay on the part of the bidder/Indian Agent would be viewed seriously and he shall be directly responsible for any loss sustained by the purchaser on the event of the delay.

2.15 PACKING

- 2.15.1 The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination as indicated in the Contract. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.
- 2.15.2 The packing, marking and documentation within and outside the packages shall comply strictly with such special requirements as shall be provided for in the Contract including additional requirements, if any, specified in SCC and in any subsequent instructions ordered by the Purchaser

2.16 Delivery and Documents

- 2.16.1 Delivery of the Goods and completion and related services shall be made by the Supplier in accordance with the terms specified by the Purchaser in the contract. The details of shipping and/or other documents to be furnished by the supplier are specified in SCC.
- 2.16.2 The terms FOB, FCA, CIF, CIP etc shall be governed by the rules prescribed in the current edition of the Incoterms published by the International Chambers of Commerce, Paris.
- 2.16.3 The mode of transportation shall be as specified in SCC

2.17 INSURANCE

- 2.17.1 Should the purchaser elect to buy on CIF/CIP basis, the Goods supplied under the Contract shall be fully insured in Indian Rupees against any loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the manner specified in SCC.
- 2.17.2 Where delivery of the goods is required by the purchaser on CIF or CIP basis the supplier shall arrange and pay for Cargo Insurance, naming the purchaser as beneficiary and initiate & pursue claims till settlement, on the event of any loss or damage.
- 2.17.3 Where delivery is on FOB or FCA basis, insurance would be the responsibility of the purchaser.
- 2.17.4 With a view to ensure that claims on insurance companies, if any, are lodged in time, the bidders and /or the Indian agent shall be responsible for follow up with their principals for ascertaining the dispatch details and informing the same to the Purchaser and he shall also liaise with the Purchaser to ascertain the arrival of the consignment after clearance so that immediately thereafter in his presence the consignment could be opened and the insurance claim be lodged, if required, without any loss of time. Any delay on the part of the bidder/Indian Agent would be viewed seriously and he shall be directly responsible for any loss sustained by the purchaser on the event of the delay.

2.18 TRANSPORTATION

- 2.18.1 Where the Supplier is required under the Contract to deliver the Goods FOB, transport of the Goods, up to and including the point of putting the Goods on board the vessel at the specified port of loading, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract price. Where the Supplier is required under the Contract to deliver the Goods FCA, transport of the Goods and delivery into the custody of the carrier at the place named by the Purchaser or other agreed point shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract price.
- 2.18.2 Where the Supplier is required under the Contract to deliver the Goods CIF or CIP, transport of the Goods to the port of destination or such other named place of destination in the Purchaser's country, as shall be specified in the Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.
- 2.18.3 In the case of supplies from within India, where the Supplier is required under the Contract to transport the Goods to a specified destination in India, defined as the Final Destination, transport to such destination, including insurance and storage, as specified in the Contract, shall be arranged by the Supplier, and the related costs shall be included in the Contract Price.

2.19 INCIDENTAL SERVICES

2.19.1 The supplier may be required to provide any or all of the services, if any, specified in SCC

2.20 SPARE PARTS

- 2.20.1 The Supplier shall be required to provide any or all of the following materials, notifications,
 - and information pertaining to spare parts manufactured or distributed by the Supplier:
 - (a) Such spare parts as the Purchaser may elect to purchase from the Supplier, providing that this election shall not relieve the Supplier of any warranty obligations under the Contract; and
 - (b) In the event of termination of production of the spare parts:
 - (i) Advance notification to the Purchaser of the pending termination, in sufficient time to permit the Purchaser to procure needed requirements; and

(ii) Following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and specifications of the spare parts, if requested.

2.21 WARRANTY

- 2.21.1 The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.
- 2.21.2 The Supplier further warrants that the Goods shall be free from defects arising from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in India.
- 2.21.3 Unless otherwise specified in the SCC, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the SCC, or for eighteen (18) months after the date of shipment from the port or place of loading in the country of origin, whichever period concludes earlier.
- 2.21.4 The Purchaser shall give notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof.

 The Purchaser shall afford all reasonable opportunity for the Supplier to inspect

such defects.

- 2.21.5 Upon receipt of such notice, the Supplier shall, within a reasonable period of time, expeditiously repair or replace the defective Goods or parts thereof, at no cost to the Purchaser.
- 2.21.6 If having been notified, the Supplier fails to remedy the defect within the reasonable period of time; the Purchaser may proceed to take within a reasonable period such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Purchaser may have against the Supplier under the Contract.
- 2.21.7 Goods requiring warranty replacements must be replaced on free of cost basis to the purchaser.

2.22 TERMS OF PAYMENT

- 2.22.1 The method and conditions of payment to be made to the Supplier under this Contract shall be as specified in the SCC.
- 2.22.2 The Supplier's request(s) for payment shall be made to the Purchaser in writing, accompanied by an invoice describing, as appropriate, the Goods delivered and the Services performed, and by documents, submitted pursuant to Delivery and document Clause of the GCC and upon fulfillment of other obligations stipulated in the contract.
- 2.22.3 Payments shall be made promptly by the Purchaser but in no case later than thirty (30) days after submission of the invoice or claim by the Supplier.
- 2.22.4 Payment shall be made in currency as indicated in the contract.

2.23. CHANGE ORDERS AND CONTRACT AMENDMENTS.

- 2.23.1 The Purchaser may at any time, by written order given to the Supplier pursuant to Clause on Notices of the GCC make changes within the general scope of the Contract in any one or more of the following:
 - (a) Drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Purchaser;
 - (b) The method of shipping or packing;
 - (c) The place of delivery; and/or
 - (d) The Services to be provided by the Supplier.
 - (e) The delivery schedule.
- 2.23.2 If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or delivery schedule, or

- both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this clause must be asserted within fifteen (15) days from the date of the Supplier's receipt of the Purchaser's change order.
- 2.23.3 No variation or modification in the terms of the contract shall be made except by written amendment signed by the parties.

2.24. ASSIGNMENT

2.24.1 The Supplier shall not assign, in whole or in part, its obligations to perform under the Contract, except with the Purchaser's prior written consent.

2.25. SUBCONTRACTS

2.25.1 The Supplier shall notify the Purchaser in writing of all subcontracts awarded under this Contract if not already specified in the bid. Such notification, in the original bid or later, shall not relieve the Supplier from any liability or duties or obligation under the Contract.

2.26 EXTENSION OF TIME.

- 2.26.1 Delivery of the Goods and performance of the Services shall be made by the Supplier in accordance with the time schedule specified by the Purchaser.
- 2.26.2 If at any time during performance of the Contract, the Supplier or its sub-contractor(s) should encounter conditions impeding timely delivery of the Goods and performance of Services, the Supplier shall promptly notify the Purchaser in writing of the fact of the delay, it's likely duration and its cause(s). As soon as practicable after receipt of the Supplier's notice, the Purchaser shall evaluate the situation and may, at its discretion, extend the Supplier's time for performance with or without penalty, in which case the extension shall be ratified by the parties by amendment of the Contract.
- 2.26.3 Except as provided under the Force Majeure clause of the GCC, a delay by the Supplier in performance of its delivery obligations shall render the Supplier liable to the imposition of penalty pursuant to Penalty Clause of the GCC unless an extension of time is agreed upon pursuant to above clause without the application of penalty clause.

2.27 PENALTY CLAUSE

2.27.1 Subject to GCC Clause on Force Majeure, if the Supplier fails to deliver any or all of the Goods or to perform the Services within the period(s) specified in the Contract, the Purchaser shall, without prejudice to its other remedies under the Contract, deduct from the Contract Price, as penalty, a sum equivalent to the percentage specified in SCC of the delivered price of the delayed Goods or unperformed Services or contract value for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the Percentage specified in SCC. Once the maximum is reached, the Purchaser may consider termination of the Contract pursuant to GCC. Clause on Termination for Default. The SCC shall also indicate the basis for ascertaining the value on which the penalty shall be applicable

2.28. TERMINATION FOR DEFAULT

- 2.28.1 The Purchaser may, without prejudice to any other remedy for breach of contract, by written notice of default sent to the Supplier, terminate the Contract in whole or part
 - (a) If the Supplier fails to deliver any or all of the Goods within the period(s) specified in the contract, or within any extension thereof granted by the Purchaser pursuant to GCC Clause on Extension of Time; or
 - (b) If the Supplier fails to perform any other obligation(s) under the Contract.
 - (c) If the Supplier, in the judgment of the Purchaser has engaged in corrupt or fraudulent or collusive or coercive practices as defined in GCC Clause on Fraud or Corruption in competing for or in executing the Contract.
- 2.28.2 In the event the purchaser terminates the contract in whole or in part, he may take recourse to any one or more of the following action:

- (a) The Performance Security is to be forfeited;
- (b) The purchaser may procure, upon such terms and in such manner as it deems appropriate, stores similar to those undelivered, and the supplier shall be liable for all available actions against it in terms of the contract.
- (c) However, the supplier shall continue to perform the contract to the extent not terminated.

2.29 FORCE MAJEURE

- 2.29.1 Notwithstanding the provisions of GCC Clauses relating to extension of time, penalty and Termination for Default the Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
- 2.29.2 For purposes of this Clause, "Force Majeure" means an event or situation beyond the control of the Supplier that is not foreseeable, is unavoidable, and its origin is not due to negligence or lack of care on the part of the Supplier. Such events may include, but not be limited to, acts of the Purchaser in its sovereign capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions, and freight embargoes.
- 2.29.3 If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof within 21 days of its occurrence. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- 2.29.4 If the performance in whole or in part or any obligations under the contract is prevented or delayed by any reason of force majeure for a period exceeding 60 days, either party may at its option terminate the contract without any financial repercussions on either side.

2.30 TERMINATION FOR INSOLVENCY

2.30.1 The Purchaser may at any time terminate the Contract by giving written notice to the Supplier, if the Supplier becomes bankrupt or otherwise insolvent. In this event, termination will be without compensation to the Supplier, provided that such termination will not prejudice or affect any right of action or remedy, which has accrued or will accrue thereafter to the Purchaser.

2.31 TERMINATION FOR CONVENIENCE

- 2.31.1 The Purchaser, by written notice sent to the Supplier, may terminate the Contract, in whole or in part, at any time. The notice of termination shall specify that termination is for the Purchaser's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date upon which such termination becomes effective.
- 2.31.2 The Goods those are complete and ready for shipment within 30 days after the Supplier's receipt of notice of termination shall be accepted by the Purchaser at the Contract terms and prices. For the remaining Goods, the Purchaser may elect:
 - (a) To have any portion completed and delivered at the Contract terms and prices; and/or
 - (b) To cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and for materials and parts previously procured by the Supplier.

2.32 SETTLEMENT OF DISPUTES

2.32.1 The Purchaser and the supplier shall make every effort to resolve amicably by direct informal negotiation any disagreement or dispute arising between them under or in connection with the Contract.

- 2.32.2 If, after twenty-one (21) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the Purchaser or the Supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided, as to the matter in dispute, and no arbitration in respect of this matter may be commenced unless such notice is given. Any dispute or difference in respect of which a notice of intention to commence arbitration has been given in accordance with this Clause shall be finally settled by arbitration. Arbitration may be commenced prior to or after delivery of the Goods under the Contract.
- 2.32.3 The dispute settlement mechanism/arbitration proceedings shall be concluded as under:
 - (a) In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any matter arising out of or connected with this agreement, such disputes or difference shall be settled in accordance with the Indian Arbitration & Conciliation Act, 1996, the rules there under and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The dispute shall be referred to the Chairman, Board of Governors of IISER TVM and if he is unable or unwilling to act, to the sole arbitration of some other person appointed by him willing to act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all parties to this order.
 - (b) In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled by arbitration In accordance with provision of subclause (a) above. But if this is not acceptable to the supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United Nations Commission on International Trade Law) Arbitration Rules.
- 2.32.4 The venue of the arbitration shall be the place from where the purchase order or contract is issued.
- 2.32.5 Notwithstanding any reference to arbitration herein,
 - (a) the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree; and
 - (b) the Purchaser shall pay the Supplier any monies due the Supplier.

2.33 GOVERNING LANGUAGE

2.33.1 The contract shall be written in English language which shall govern its interpretation. All correspondence and other documents pertaining to the Contract, which are exchanged by the parties, shall be written in the English language only.

2.34 APPLICABLE LAW

2.34.1 The Contract shall be interpreted in accordance with the laws of the Union of India and all disputes shall be subject to place of jurisdiction as specified in SCC.

2.35 Notices

- 2.35.1 Any notice given by one party to the other pursuant to this contract/order shall be sent to the other party in writing or by cable, telex, FAX, e-mail or and confirmed in writing to the other party's address specified in the SCC.
- 2.35.2 A notice shall be effective when delivered or on the notice's effective date, whichever is later.

2.36 TAXES AND DUTIES

- 2.36.1 For goods manufactured outside India, the Supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside India.
- 2.36.2 For goods Manufactured within India, the Supplier shall be entirely responsible

for all taxes, duties, license fees, etc., incurred till its final manufacture/production.

2.36.3 If any tax exemptions, reductions, allowances or privileges may be available to the Supplier in India, the Purchaser shall make its best efforts to enable the Supplier to benefit from any such tax savings to the maximum allowable extent.

2.37 RIGHT TO USE DEFECTIVE GOODS

2.37.1 If after delivery, acceptance and installation and within the guarantee and warranty period,

the operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or complete replacement is made without interfering with the Purchaser's operation.

2.38 PROTECTION AGAINST DAMAGE

2.38.1 The system shall not be prone to damage during power failures and trip outs. The normal voltage and frequency conditions available at site as under:

- (a) Voltage 230 volts Single phase / 415 V 3 phase (+_ 10%)
- (b) Frequency 50 Hz.

2.39. SITE PREPARATION AND INSTALLATION

The Purchaser is solely responsible for the construction of the equipment sites in compliance with the technical and environmental specifications defined by the Supplier. The Purchaser will designate the installation sites before the scheduled installation date to allow the Supplier to perform a site inspection to verify the appropriateness of the sites before the installation of the Equipment, if required. The supplier shall inform the purchaser about the site preparation, if any, needed for installation, of the goods at the purchasers' site immediately after notification of award/contract.

Sd/-

DEPUTY REGISTRAR [F&A]
ADDITIONAL CHARGE [P&S]
IISERTVM

INTEGRITY PACT

Between

Indian Institute of Science Education and Research, Thiruvananthapuram (IISER-TVM) hereinafter referred to as "The Principal".

And

herein referred to as "The Bidder/ Contractor."

Preamble

The Principal will nominate an Independent External Monitor (IEM) by name at the tender stage/will appoint in case of receipt of any reference, from the panel of IEMs, for monitoring the tender process and/or the execution of the contract in order to ensure compliance with the Integrity Pact by all the parties concerned.

<u>Section 1 - Commitments of the Principal / Purchaser / Employer:</u>

- (1) The Principal commits itself to take all measures necessary to prevent corruption and to observe the following principles:
 - a) No employee of the Principal, personally or through family members, will in connection with the tender for, or the execution of a contract, demand, takes a promise for or accepts, for self or third person, any material or immaterial benefit which the person is not legally entitled to.
 - b) The Principal will, during the tender process treat all Bidder(s) with equity and reason. The Principal will in particular, before and during the tender process, provide to all Bidder(s) the same information and will not provide to any Bidder(s) confidential/additional information through which the Bidder(s) could obtain an advantage in relation to the tender process or the contract execution.
 - c) The Principal will exclude from the process all known prejudiced persons.
- (2) If the Principal obtains information on the conduct of any of its employees which is a criminal offence under the IPC/PC Act, or if there be a substantive suspicion in this regard, the Principal will inform the Chief Vigilance Officer of IISER-TVM and in addition can initiate disciplinary action.

<u>Section 2 - Commitments of the Bidder(s)/Contractor(s)</u>

- (1) The Bidder(s)/Contractor(s) commit himself to take all measures necessary to prevent corruption. He commits himself to observe the following principles during his participation in the tender process and during the contract execution:
 - a. The Bidder(s)/Contractor(s) will not, directly or through any other person or firm, offer, promise or give to any of the Principal's employees involved in the tender process or the execution of the

- contract or to any third person any material or other benefit which he/she is not legally entitled to, in order to obtain in exchange any advantage of any kind whatsoever during the tender process or during the execution of the contract or to vitiate the Principal's tender process or contract execution.
- b. The Bidder(s)/Contractor(s) will not enter with other Bidders into any undisclosed agreement or understanding, whether formal or informal. This applies in particular to prices, specifications, certifications, subsidiary contracts, submission or non-submission of bids or any other actions to restrict competitiveness or to introduce cartelization in the bidding process or to vitiate the Principal's tender process or execution of the contract.
- c. The Bidder(s)/Contractor(s) will not commit any offence under the relevant IPC/PC Act; further the Bidder(s)/Contractor(s) will not use improperly, for purposes of competition or personal gain, or pass on to others, any information or document provided by the Principal as part of the business relationship, regarding plans, technical proposals and business details, including information contained or transmitted electronically.
- d. The Bidder(s)/Contractor(s) of foreign origin shall disclose the name and address of the Agents/representatives in India, if any. Similarly the Bidder(s)//Contractors(s) of Indian Nationality shall furnish the name and address of the foreign supplier / contract agency, if any. Further details as mentioned in the "Guidelines on Indian Agents of Foreign Suppliers" / Contract Agencies", shall be disclosed by the Bidder(s)/Contractor(s) wherever applicable. Further, as mentioned in the Guidelines all the payments made to the Indian agent (s)/representative (s) have to be in Indian Rupees only. Copy of the "Guidelines on Indian Agents of Foreign Suppliers/ Contract Agencies" is annexed and marked as Annexure A.
- e. The Bidder(s)/Contractor(s) will, when presenting his bid, disclose any and all payments he has made or committed to or intends to make to agents, brokers or any other intermediaries in connection with the award of the contract.
- (2) The Bidder(s)/Contractor(s) will not instigate third persons to commit offences outlined above or be an accessory to such offences.

<u>Section 3 - Disqualification from tender process and exclusion from future contracts.</u>

- (1) A transgression is considered to have occurred, if the Principal after due consideration of the available evidence, concludes that reasonable doubt is possible.
- (2) If the Bidder(s)/Contractor(s), before award of contract or after award of Contract or during execution has committed a transgression through a

violation of Section 2 above or in any other form such as to put his reliability or credibility in question, the Principal is entitled to disqualify the Bidder(s)/Contractor(s) from the tender process or to terminate the Contract, if already awarded or take action as per the procedure mentioned in the "Guidelines on Banning of business dealings". Copy of the "Guidelines on Banning of business dealings" is annexed and marked as Annexure - B

- (3) If the Bidder (s) / Contractor (s) has committed transgression through a violation of any of the terms under Section 2 above or in any other form such as to put his reliability or credibility into question, the Principal is also exclude bidder/contractor entitled to the from tenders/contract award processes. The imposition and duration of the exclusion will be determined by the Principal keeping in view the severity The severity will be determined transgression. circumstances of the case, in particular the number of transgression, the position of the transgressors within the company hierarchy of the bidder/contractor and the amount of the damage.
- (4) If it is observed after payment of final bill but before the expiry of validity of Integrity Pact that the contractor has committed a transgression through a violation of any of the terms under Section 2 above during the execution of contract, the Principal is entitled to exclude the contractor from future tenders / contract award processes.
- (5) The exclusion will be imposed for a minimum period of six (6) months and a maximum period of 3 (three) years.
- (6) If the Bidder/Contractor can prove that he has restored/recouped the damage to the Principal caused by him and has installed a suitable corruption prevention system, the Principal may revoke the exclusion before the expiry of the period of such exclusion.

<u>Section</u> <u>4 - Compensation for Damages</u>

- (1) If the Principal has disqualified the Bidder(s) from the tender process prior to the award in accordance with Section 3 above, the Earnest Money Deposit (EMD)/Bid Security furnished, if any, along with the offer as per the terms of the Invitation to Tender (ITT) shall be forfeited. This is apart from the disqualification of the Bidder as may be imposed by the Principal as brought out at Section 3 above.
- (2) If the Principal has terminated the contract according to Section 3, or if the Principal is entitled to terminate the contract according to Section 3, the Principal shall be entitled to forfeit the Security Deposit/Performance Bank Guarantee furnished by the Bidder/Contractor or to demand and recover from the Contractor liquidated damages of the contract value or the amount equivalent to Performance Bank Guarantee. This is apart from the disqualification of the Bidder as may be imposed by the Principal as brought out at Section 3 above.

Section 5 - Previous transgression

- (1) The Bidder/Contractor declares that no previous transgressions occurred in the last 3 years with any other Company in any country conforming to the anti-corruption approach or with Government/any other Public Sector Enterprise in India that could justify his exclusion from the tender process /contract.
- (2) If the Bidder/Contractor makes incorrect statement on this subject, he can be disqualified from the tender process /contract either it can be terminated for such reason or action can be taken as per the procedure mentioned in "Guidelines on Banning of business dealings."

<u>Section 6 - Equal treatment of all Bidders / Contractors/ Subcontractors.</u>

- (1) The Bidder(s)/Contractor(s) undertake(s) to demand from all his Subcontractors a commitment in conformity with this Integrity Pact, and to submit it to the Principal before seeking permission for such sub-contract signing.
- (2) The Principal will enter into agreements with identical conditions as this one with all Bidders and Contractors.
- (3) The Principal will disqualify from the tender process all bidders who do not sign this Pact or violate its provisions.

<u>Section 7 - Criminal charges against violating Bidders / Contractors/</u> Subcontractors.

If the Principal obtains knowledge of conduct of a Bidder, Contractor, Sub-contractor or of any employee or a representative or an associate of a Bidder/Contractor/Sub-contractor which constitutes corruption, or if the Principal has substantive suspicion in this regard, the Principal will inform the same to Chief Vigilance Officer.

<u>Section 8 - Independent External Monitor (s) (IEMs)</u>

- (1) The Principal appoints competent and credible Independent External Monitor (s) (IEMs) with clearance from Central Vigilance Commission. IEMs reviews independently, the cases referred to him/them to assess whether and to what extent the parties concerned comply with the obligations under this Integrity Pact.
- (2) In case of non-compliance of the provisions of the Integrity Pact, the complaint/non-compliance is to be lodged by the aggrieved party with the Nodal Officer only appointed by Director, IISER-TVM. The Nodal Officer shall refer the complaint/non-compliance so received by him to the IEM, already appointed or to be appointed for that case.
- (3) The Monitor is not subject to instructions by the representatives of the parties and performs his functions neutrally and independently. He reports

to the Director, IISER-TVM.

- (4) The Bidder(s)/Contractor(s) accepts that the Monitor has the right to access without restriction, to all tender/contract documentation of the Principal including that provided by the Bidder/Contractor. The Bidder/Contractor will also grant the Monitor, upon his request and demonstration of a valid interest, unrestricted and unconditional access to his tender/contract documentation. The same is applicable to Sub-contractors also. The Monitor is under contractual obligation to treat the information and documents of the Bidder(s)/Contractor(s)/Sub-contractor(s) with confidentiality.
- (5) The Principal will provide to the Monitor sufficient information about all meetings among the parties related to the tender/contract provided such meetings could have an impact on the contractual relations between the Principal and the Contractor. The parties offer to the Monitor the option to participate in such meetings.
- (6) As soon as the Monitor notice, or believes to notice, a violation of this Integrity Pact, he will so inform the Management of the Principal and request the Management to discontinue or take corrective action, or to take other relevant action. The monitor can in this regard submit non-binding recommendations. Beyond this, the Monitor has no right to demand from the parties that they act in a specific manner, refrain from action or tolerate action.
- (7) The Monitor will submit a written report to the Director, IISER-TVM within 8 to 10 weeks from the date of reference or intimation to him by the Principal and should the occasion arise, submit proposals for correcting problematic situations for the violations or the breaches of the provisions of the agreement noticed by the Monitor.
- (8) Monitor shall be entitled to compensation on the same terms as being extended to/provided to Director IISER-TVM.
- (9) If the Monitor has reported to the Director, IISER-TVM, a substantiated suspicion of an offence under relevant IPC/PC Act, and the Director, IISER-TVM has not, within the reasonable time taken visible action to proceed against such offence or reported it to the Chief Vigilance Officer, the Monitor may also transmit this information directly to the Central Vigilance Commissioner.
- (10) The word 'Monitor' means Independent External Monitor and would include both singular and plural.

Section 9 - Pact Duration

- (1) This Pact begins when both parties have legally singed it. It expires for the Contractor 12 months after the last payment under the contract, and for all other unsuccessful Bidders 6 months after the contract has been awarded.
- (2) If any claim is made/lodged during the valid period of the Integrity Pact,

the same shall be binding and continue to be valid despite the lapse of this pact as specified above, unless it is discharged/determined by Director, IISER-TVM.

Section 10 - Other provisions

- (1) This agreement is subject to Indian Law. Place of performance and Jurisdiction is Thiruvananthapuram, Kerala.
- (2) Changes and supplements as well as termination notices need to be made in writing. Side agreements to this Integrity Pact have not been made.
- (3) If the Contractor is a partnership or a consortium, this agreement must be signed by all partners or consortium members or their Authorised Representative (s) by duly furnishing Authorisation to sign Integrity Pact.
- (4) Should one or several provisions of this agreement turn out to be invalid, the remaining part of this agreement remains valid. In this case, the parties will strive to come to an agreement to their original intentions.
- (5) Wherever he or his is indicated in the above sections, the same may be read as he/she or his/her, as the case may be.

(For & On behalf of the Principal)	(For & On behalf of Bidder/Contractor)
(Office Seal)	(Office Seal)
Place	Place Date
Date	
Witness 1:	Witness 1:
(Name & Address):	(Name & Address):
Witness 2:	Witness 2:
(Name & Address):	(Name & Address):

MODEL BANK GUARANTEE FORMAT FOR FURNISHING BID SECURITY (BS)

Whereas	(hereinafte	called	the "tende	erer")
has submitted their offer dated		or the su	pply of	
	(h	ereinafter	called	the
"tender") against the Purchaser's tender	enquiry N	o		
KNOW ALL MEN by these prese	ents that V	/E		
of having or	ur register	ed office	at	
are bound unto(hereinafter	called the	"Purchase	er) in
the sum of			for w	vhich
payment will and truly to be made to the said l	Purchaser,	he Bank	binds itsel	lf, its
successors and assigns by these presents. Se	ealed with t	he Comm	on Seal o	f the
said Bank thisday of20				

THE CONDITIONS OF THIS OBLIGATION ARE:

- 1) If the tenderer withdraws or amends, impairs or derogates from the tender in any respect within the period of validity of this tender.
- 2) If the tenderer having been notified of the acceptance of his tender by the Purchaser during the period of its validity:
 - a) If the tenderer fails to furnish the performance security for the due performance of the contract.
 - b) Fails or refuses to accept/execute the contract.

We undertake to pay the Purchaser up to the above amount upon receipt of its first written demand, without the purchaser (IISER-TVM) having to substantiate its demand, provided that in its demand the purchaser will note that the amount claimed by it is due to it owing the occurrence of one or both the two conditions, specifying the occurred condition or conditions. This guarantee will remain in force up to and including 45 days after the period of tender validity and any demand in respect thereof should reach the bank not later than the above date.

Signature of the authorized officer of the bank

Name and designation of the officer

Seal, name and address of the Bank and address of the Branch.