



**INDIAN INSTITUTE OF SCIENCE EDUCATION AND
RESEARCH THIRUVANANTHAPURAM [IISERTVM]**

PH.-0471 2778019,
EMAIL: purchasestores@iisertvm.ac.in

MARUTHAMALA.P.O VITHURA. P.O
THIRUVANANTHAPURAM 695551,
KERALA, INDIA
GST No.32AAAJI0299R1ZS

IISER/PUR/0295/AP/FAA/SEESS /25-26

Date: 8.08.2025

CORRIGENDUM

No: IISER/PUR/0295/AP/FAA/SEESS /25-26

Sub: Supply, Installation, and commissioning of Inductively Coupled Plasma Mass Spectrometer (ICP-MS) with Microwave Digestion System: **reg**

Ref: Tender ID: 2025_IISRT_866523_1

1. The technical specification, point no 7 of the above mentioned tender is revised as follows :

Sl No	In Place of	To Read as
#6 Mass analyser (Quadrupole): point no. 7	- Abundance sensitivity should be less than or equal to 5×10^{-7} (low-mass side) and less than or equal to 1×10^{-7} (high-mass side)..	Abundance Sensitivity should be less than or equal to 5×10^{-7} (low mass side) & less than or equal to 5×10^{-7} (high mass side)'.

2. All other Terms and Conditions remain the same. Bidders may quote accordingly

Thanking You,

Yours Faithfully

Amman Sadath
8/8/25
Assistant Registrar (P&S)